

Specifications for Lab Equipment 2025-2026 Budget

Specifications for VITEK 2 Compact 60, Culture & Sensitivity Machine

No.	Item	Specifications
1	VITEK 2 Compact 60 Culture & Sensitivity Machine	Fully auto organism identification and antibiotic sensitivity machine Capacity to accommodate a minimum of 60 tests (Either 60 ID and/or AST tests), at any time. Incubation temperature : 35 degree to ambient + 5° Shaking cycle : 30 cycle /min or more Dimensions : 60 cm (H)x 72 cm(W)x 68cm (D) Self-quality control and calibration system : Possible Display : External LCD monitor Printer : External

Specifications for Lab Equipment 1 Lot: Micropipette & Magnetic Hotplate Stirrer

No.	Item	Specifications
1	5ml Micropipette adjustable single channel pipette	Brand- Eppendorf, Germany, Single Channel, Variable 0.5-5mL
2	10ml Micropipette adjustable single channel pipette	Brand- Eppendorf, Germany, Single Channel, Variable 1-10mL
3	Magnetic Hotplate Stirrer	Preferable Brand: IKA, Germany Stir volume liters: Up to 5 Liters Switch on button with indicator lamp for heating and for stirring speed. Electronic stir speed control. Electronic heater energy control from 0-100%,