

Specifications for Lab Equipment 2025-2026 Budget

Specifications for VITEK 2 Compact 60, Culture & Sensitivity Machine

No.	Item	Specifications
1	VITEK 2 Compact 60 Culture & Sensitivity Machine	Fully auto organism identification and antibiotic sensitivity machine Capacity to accommodate a minimum of 60 tests (Either 60 ID and/or AST tests), at any time. Incubation temperature : 35 degree to ambient + 5° Shaking cycle : 30 cycle /min or more Dimensions : 60 cm (H)x 72 cm(W)x 68cm (D) Self-quality control and calibration system : Possible Display : External LCD monitor Printer : External