

စိုက်ပျိုးရေး၊ မွေးမြူရေးနှင့် ဆည်မြောင်းဝန်ကြီးဌာန၊
မွေးမြူရေးဆိုင်ရာဆေးတက္ကသိုလ်

Lot 12 Lab Equipment Specifications

No.	Item	Specifications	Quantity
1.	Compact centrifuge	<ul style="list-style-type: none"> - Temperature range: -10°C to + 40°C - Capacity: 15 ml and 50 ml centrifuge tubes - Maximum Speed: 10,000 rpm for 50 ml, 12,000 rpm for 15 m - Run Time: 99 hr 59 min, continuous - Type: bench or floor - Power supply: AC 230 V, 50- 60 Hz - Accessories: rotors (for 15 ml tube and 50 ml) - Must be included accessories. 	1 No
2	Autoclave	<ul style="list-style-type: none"> - Capacity: 50L / Type: Vertical - Working pressure: 0.23MPa - Working Temperature: 105°C - 136°C - Temperature Precision: 0.1°C - Power supply: AC 220V, 50/60Hz 	1 No
3	Multifunctional water bath shaker	<ul style="list-style-type: none"> - Mode: Reciprocating oscillation + thermostatic water bath - Temperature range: 20°C - 100°C - Temperature Precision: 0.1°C - Reciprocating Frequency range: 20 - 180 rpm - Water tank: stainless steel - Timer: 0 - 9999 min - Power supply: AC 220V, 50/60Hz - Accessories: Water bath flask seat, test tube rack - Must be included accessories. 	1 No
4	Deep Freezer	<ul style="list-style-type: none"> - Type: Chest - Temperature range: -5°C to - 40°C - Capacity: minimum 200L - Power supply: AC 220V, 50/60Hz - Accessories: power safe guard - Must be included accessories. 	1 No