

စိုက်ပျိုးရေး၊ မွေးမြူရေးနှင့် ဆည်မြောင်းဝန်ကြီးဌာန၊
 မွေးမြူရေးဆိုင်ရာဆေးတက္ကသိုလ်

Lot 8 Lab Equipment Specifications

No.	Item	Specifications	Quantity
1.	Semi- Automatic Microtome with blades set	- Section Thickness Setting - 0.5-100 microns - Trimming Thickness- 1-600 microns - Object Feed- 28mm - Vertical Stroke- 60mm - Max. Specimen Size - 50 x 50 mm - Specimen Clamp Orientation- 360 degrees - Nominal Voltage - 100/120/230/240V AC +/- 10% - Power draw – 200VA - Dimensions – 580mm×360mm×310mm - Weight – 35kgs	1 No
2	Digital Weighing Scale	- Maximum weight – 30kg - Size – 350mm× 450mm	1 No